



KENTUCKY COMMUNITY AND TECHNICAL COLLEGE SYSTEM

INVITATION TO BID ADDENDUM

SOLICITATION NO.: KCT-01392

ADDENDUM NO.: 1

SOLICITATION ISSUE DATE: August 21, 2023

ADDENDUM DATE: August 24, 2023

OPENING DATE: AUGUST 30, 2023, 4:00PM EST

TITLE: Radiographic Overhead Tube Crane System w/LCD Touchscreen & Elevating Table

The following information is being provided in response to questions received for this BID:

- 1.Q) Is the site prepared with the proper Unistrut in the ceiling and at what height?
a) The ceiling has an existing Unistrut for a separate unit. Ceiling height estimated 9'.
- 2.Q) Is power installed in the room and what is the specification for the power?
a) The room has single phase.
- 3.Q) Page 11 under P/N specifies single phase 208-240V and on page 13, line 2 states required three-phase electrical power. Please clarify.
a) Page 11 is correct, requested power is single phase. Please disregard Line #2 and use the Form of Proposal to provide your pricing for this bid.

Line	Item #/ Description	Qty	Extended Cost
1	QTY -1 CARESTREAM 40KW w/AEC Q-VISION/LCD Touchscreen OTC System As specified in Section 10 of the bid.	1	\$
2	Please disregard Line #2 and use the Form of Proposal to provide your pricing for this bid.		
3	Freight/shipping and Installation charges (must include all costs for delivery and install, no hidden fees)– Delivery to Jefferson CTC, 110 W. Chestnut, Louisville KY 40202. Loading dock accessible 110 W Chestnut / 1 st Street. Forklift not required by supplier.		\$
GRAND TOTAL			\$
Warranty: _____ Can guarantee delivery and complete installation on/before December 15, 2023: _____ **DUE TO GRANT FUNDING, THIS MUST BE DELIVERED and INSTALLED NO LATER THAN DECEMBER 15, 2023			

The time to submit questions for this Request for Proposal has now expired. No further questions will be accepted.

All other terms, conditions & specifications remain unchanged.

Offerors must acknowledge receipt of this and any addenda either with solicitation or by separate letter or email prior to award of contract. If by separate letter, the following information should be placed in the lower left-hand corner of the envelope:

SOLICITAION NO.: KCT-01392
TITLE: Radiographic Overhead Tube Crane System w/LCD Touchscreen & Elevating Table

Name of Firm: _____

Authorized Signature: _____